
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002059
14:05 19/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:42
STERILIZING 3:38
136.0 C 220kPa 3:38
136.1 C 223kPa 4:08
136.2 C 223kPa 4:38
136.4 C 225kPa 5:08
136.2 C 223kPa 5:38
136.4 C 225kPa 6:08
136.3 C 224kPa 6:38
136.2 C 223kPa 7:08
136.2 C 223kPa 7:08
VENTING 7:08
AIR DRYING 8:00
CYCLE COMPLETE 8:15

Digital Signature #
5CE57E55753466CB
