

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2969

Name: RITTER 1  
Sterilizer ID: V2178198  
03/03/2023 10:45 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 181.0 F 2.7  
4:00 191.4 F 8.4  
6:00 208.5 F 6.4  
8:00 232.0 F 9.9  
10:00 233.5 F 7.9  
12:00 239.2 F 9.5  
14:00 250.1 F 15.0  
16:00 263.9 F 22.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 271.8 F 27.9  
2:00 272.5 F 28.3  
3:00 272.8 F 28.8  
  
Min 271.4 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
Dry Cycle Aborted

FILLING: 1:10  
HEATING: 17:46  
STERILIZING: 3:00  
VENTING: 2:47  
DRYING: 0:08  
TOTAL CYCLE: 00:24:51

03/03/2023 11:09 AM  
ACCEPTED:  
REJECTED:  
Reason: -----  
Operator: -----