
Midmark M11 - v1.0.5
Total Cycles: 3452

Name: RITTER2
Sterilizer ID: V2313228
03/12/2024 12:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.5 F 0.0
2:00 106.3 F 0.0
4:00 116.0 F 0.1
6:00 152.5 F 2.8
8:00 176.5 F 6.1
10:00 190.6 F 10.0
12:00 209.4 F 8.4
14:00 222.5 F 7.9
16:00 233.5 F 12.3
18:00 234.4 F 9.5
20:00 238.1 F 10.2
22:00 246.2 F 14.0
24:00 257.7 F 20.2
26:00 268.1 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.9 F 30.0
2:00 272.2 F 29.5
3:00 271.0 F 29.1
4:00 271.8 F 29.9

Min 270.9 F 29.0
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 213.7 F 0.0
10:00 210.1 F 0.0
15:00 204.9 F 0.0
20:00 203.3 F 0.0
25:00 203.3 F 0.0
30:00 202.4 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 27:22
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:05:58
CYCLE COMPLETE

03/12/2024 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____