
Midmark M11 - v1.0.0
Total Cycles: 874

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2021 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.6 F 0.0
2:00 130.2 F 0.8
4:00 150.4 F 2.9
6:00 180.1 F 6.4
8:00 211.5 F 11.6
10:00 218.3 F 5.3
12:00 227.9 F 10.0
14:00 232.7 F 8.6
16:00 242.3 F 12.4
18:00 240.2 F 10.6
20:00 253.2 F 17.1
22:00 264.0 F 23.7
24:00 271.1 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 273.6 F 29.4
2:00 271.5 F 28.6
3:00 272.7 F 29.5
4:00 272.6 F 29.0

Min 271.1 F 28.5
Max 273.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 243.8 F 0.0
10:00 239.9 F 0.0
15:00 227.7 F 0.0
20:00 223.2 F 0.0
25:00 220.2 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:16
CYCLE COMPLETE

06/08/2021 03:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____