
Midmark M11 - v1.0.0
Total Cycles: 2592

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2022 09:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	126.1 F	0.0
2:00	128.0 F	0.2
4:00	137.4 F	3.0
6:00	162.8 F	6.8
8:00	208.4 F	11.6
10:00	219.2 F	5.0
12:00	231.0 F	9.7
14:00	237.6 F	8.9
16:00	239.8 F	11.4
18:00	237.4 F	9.2
20:00	251.1 F	15.9
22:00	262.9 F	22.7
24:00	272.2 F	28.6

STERILIZING

mm:ss	Degrees	PSI
0:00	271.5 F	28.4
1:00	272.6 F	29.0
2:00	271.5 F	28.4
3:00	272.7 F	29.1
4:00	271.8 F	28.4

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	212.2 F	0.0
	5:00	206.4 F	0.0
	10:00	215.1 F	0.0
	15:00	214.8 F	0.0
	20:00	216.3 F	0.0
	25:00	220.4 F	0.0
Drying Complete	30:00	222.5 F	0.0

FILLING: 0:58
HEATING: 24:12
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:22
CYCLE COMPLETE

11/22/2022 10:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____