
Midmark M11 - v1.0.5
Total Cycles: 1483

Name: RITTER2
Sterilizer ID: V2313228
09/13/2022 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.2 F 0.0
2:00 205.3 F 3.7
4:00 203.2 F 9.5
6:00 224.4 F 7.1
8:00 233.8 F 12.1
10:00 236.0 F 10.4
12:00 236.4 F 9.4
14:00 250.9 F 16.5
16:00 262.7 F 23.5
18:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.4 F 30.1
2:00 271.6 F 29.4
3:00 271.7 F 30.0
4:00 272.3 F 30.0

Min 271.2 F 29.2
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 236.9 F 0.0
10:00 227.3 F 0.0
15:00 227.9 F 0.0
20:00 228.4 F 0.0
25:00 231.5 F 0.0
30:00 203.9 F 0.0
Drying Complete

FILLING: 2:56
HEATING: 18:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

09/13/2022 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____