

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5366

Name: RITTER 1  
Sterilizer ID: V2178198  
02/26/2025 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 78.8 F 0.4  
4:00 94.0 F 1.8  
6:00 130.0 F 4.3  
8:00 155.3 F 7.2  
10:00 179.1 F 10.5  
12:00 204.5 F 5.1  
14:00 215.2 F 4.7  
16:00 222.8 F 8.0  
18:00 233.7 F 11.7  
20:00 228.9 F 6.6  
22:00 240.2 F 11.6  
24:00 236.6 F 8.8  
26:00 249.6 F 15.0  
28:00 260.6 F 21.2  
30:00 270.6 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.5  
1:00 272.1 F 28.5  
2:00 272.5 F 28.9  
3:00 272.4 F 29.1  
4:00 271.8 F 28.4  
  
Min 271.8 F 28.4  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 207.4 F 0.0  
10:00 182.5 F 0.0  
15:00 198.9 F 0.0  
20:00 209.9 F 0.0  
25:00 208.9 F 0.0  
30:00 215.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 30:44  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:09:13  
CYCLE COMPLETE

02/26/2025 10:48 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_