
Midmark M11 - v1.0.0
Total Cycles: 1604

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2022 07:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 75.0 F 0.0
4:00 79.8 F 0.0
6:00 116.4 F 0.1
8:00 151.1 F 0.4
10:00 178.0 F 3.0
12:00 196.8 F 6.7
14:00 206.3 F 10.0
16:00 217.5 F 13.5
18:00 223.3 F 7.2
20:00 233.7 F 11.3
22:00 230.8 F 7.2
24:00 243.6 F 12.8
26:00 241.2 F 10.9
28:00 253.4 F 17.0
30:00 263.7 F 23.1
32:00 272.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.3
3:00 272.6 F 28.9
4:00 271.4 F 28.2

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 219.5 F 0.0
10:00 223.2 F 0.0
15:00 229.8 F 0.0
20:00 219.8 F 0.0
25:00 229.4 F 0.0
30:00 225.5 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:13
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:10:20
CYCLE COMPLETE

02/08/2022 08:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____