
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000792
14:30 15/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:38
PRESSURIZING 6:27
STERILIZING 7:55
136.0 C 220kPa 7:55
136.1 C 217kPa 8:25
135.9 C 219kPa 8:55
136.3 C 223kPa 9:25
136.4 C 221kPa 9:55
136.1 C 220kPa 10:25
136.4 C 218kPa 10:55
136.4 C 223kPa 11:25
136.4 C 223kPa 11:25
VENTING 11:26
AIR DRYING 12:28
CYCLE COMPLETE 42:28

Digital Signature #
4A8119962C2FD303
