
Midmark M11 - v1.0.5
Total Cycles: 1910

Name: RITTER2
Sterilizer ID: V2313228
01/10/2023 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.8 F 0.0
2:00 148.7 F 1.7
4:00 157.4 F 4.9
6:00 181.1 F 9.3
8:00 208.6 F 7.2
10:00 222.0 F 8.1
12:00 233.1 F 13.4
14:00 235.9 F 10.3
16:00 234.8 F 8.3
18:00 248.7 F 15.3
20:00 260.7 F 22.1
22:00 271.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 30.0
2:00 271.9 F 29.4
3:00 271.5 F 29.6
4:00 272.5 F 30.0

Min 271.3 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 230.6 F 0.0
10:00 231.6 F 0.0
15:00 228.9 F 0.0
20:00 228.6 F 0.0
25:00 231.3 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:46
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

01/10/2023 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____