
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 007676
15:27 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:08
STERILIZING 4:07
136.4 C 221kPa 4:07
136.4 C 223kPa 4:37
136.1 C 220kPa 5:07
136.4 C 224kPa 5:37
136.2 C 220kPa 6:07
136.4 C 223kPa 6:37
136.2 C 223kPa 7:07
136.4 C 220kPa 7:37
136.4 C 220kPa 7:37
VENTING 7:37
AIR DRYING 8:38
CYCLE COMPLETE 8:49

Digital Signature #
96103D491D8E1D0D
