
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001420
11:52 15/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:40
STERILIZING 4:49
136.0 C 220kPa 4:49
136.4 C 222kPa 5:19
136.3 C 222kPa 5:49
136.4 C 222kPa 6:19
136.0 C 220kPa 6:49
136.3 C 222kPa 7:19
136.4 C 222kPa 7:49
136.3 C 222kPa 8:19
136.3 C 222kPa 8:19
VENTING 8:19
AIR DRYING 9:10
CYCLE COMPLETE 20:13

Digital Signature #
8DBC79F7311B4824
