

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1234

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2022 10:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.5 F 0.0  
2:00 169.0 F 3.2  
4:00 184.8 F 7.8  
6:00 204.0 F 13.2  
8:00 222.0 F 7.9  
10:00 235.7 F 13.3  
12:00 237.5 F 11.0  
14:00 237.3 F 9.6  
16:00 250.8 F 16.4  
18:00 262.4 F 23.2  
20:00 272.7 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 30.0  
2:00 272.0 F 29.4  
3:00 271.5 F 29.4  
4:00 272.7 F 30.1

Min 271.4 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
5:00 208.8 F 0.0  
10:00 228.0 F 0.0  
15:00 217.3 F 0.0  
20:00 202.6 F 0.0  
25:00 202.4 F 0.0  
30:00 198.8 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:27  
CYCLE COMPLETE

06/28/2022 11:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_