

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3422

Name: RITTER 1  
Sterilizer ID: V2178198  
06/21/2023 11:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.9 F 0.0  
2:00 183.0 F 3.0  
4:00 201.1 F 7.1  
6:00 215.0 F 12.4  
8:00 225.0 F 6.5  
10:00 237.6 F 11.6  
12:00 234.9 F 8.6  
14:00 241.1 F 10.3  
16:00 246.9 F 13.5  
18:00 259.2 F 20.1  
20:00 270.0 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.5 F 28.8  
2:00 271.6 F 28.0  
3:00 272.4 F 28.9  
4:00 272.2 F 28.5  
  
Min 271.4 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 226.7 F 0.0  
10:00 232.4 F 0.0  
15:00 230.8 F 0.0  
20:00 227.6 F 0.0  
25:00 228.4 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 20:57  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

06/21/2023 12:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_