

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4713

Name: RITTER 1  
Sterilizer ID: V2178198  
07/10/2024 04:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
2:00 176.7 F 3.8  
4:00 189.5 F 8.2  
6:00 215.8 F 12.1  
8:00 227.4 F 8.0  
10:00 239.9 F 13.2  
12:00 237.0 F 10.1  
14:00 237.1 F 8.2  
16:00 249.3 F 15.1  
18:00 261.5 F 21.8  
20:00 272.4 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.4 F 28.9  
2:00 271.8 F 28.7  
3:00 272.5 F 28.9  
4:00 271.5 F 28.5  
  
Min 271.5 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 212.0 F 0.0  
10:00 196.3 F 0.0  
15:00 198.3 F 0.0  
20:00 201.7 F 0.0  
25:00 230.3 F 0.0  
30:00 232.9 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 20:26  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

07/10/2024 05:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_