
Midmark M11 - v1.0.0
Total Cycles: 5204

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2024 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.8 F 0.0
2:00 180.5 F 0.3
4:00 194.9 F 4.2
6:00 213.5 F 9.0
8:00 228.6 F 10.5
10:00 235.2 F 9.1
12:00 238.7 F 9.9
14:00 244.8 F 12.6
16:00 242.9 F 11.5
18:00 256.0 F 18.2
20:00 267.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 272.3 F 28.8
2:00 272.0 F 28.2
3:00 271.7 F 28.5
4:00 272.1 F 28.5

Min 271.1 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 226.7 F 0.0
10:00 195.7 F 0.0
15:00 192.9 F 0.0
20:00 190.9 F 0.0
25:00 248.6 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 00:59:60
CYCLE COMPLETE

12/18/2024 05:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____