
Midmark M11 - v1.0.0
Total Cycles: 1064

Name: RITTER 1
Sterilizer ID: V2178198
07/28/2021 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.9 F 0.0
2:00 179.6 F 2.4
4:00 190.0 F 5.9
6:00 214.2 F 11.4
8:00 221.1 F 5.5
10:00 232.3 F 10.7
12:00 231.4 F 7.0
14:00 244.4 F 13.0
16:00 243.8 F 12.2
18:00 256.5 F 18.8
20:00 267.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 271.8 F 28.4
3:00 272.0 F 29.0
4:00 272.5 F 29.0

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.4 F 0.0
5:00 225.9 F 0.0
10:00 229.4 F 0.0
15:00 228.5 F 0.0
20:00 224.7 F 0.0
25:00 223.2 F 0.0
30:00 199.5 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

07/28/2021 10:30 AM
ACCEPTED: X
REJECTED:
Operator: U01