
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000914
9:47 07/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:18
STERILIZING 4:24
136.0 C 220kPa 4:24
136.4 C 223kPa 4:54
136.3 C 223kPa 5:24
136.3 C 222kPa 5:54
136.4 C 221kPa 6:24
136.4 C 224kPa 6:54
136.4 C 225kPa 7:24
136.4 C 220kPa 7:54
136.4 C 220kPa 7:54
VENTING 7:54
AIR DRYING 8:49
CYCLE COMPLETE 8:52

Digital Signature #
31642D974502D508
