
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001449
10:14 09/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:41
PRESSURIZING 6:22
STERILIZING 7:52
136.0 C 220kPa 7:52
136.0 C 217kPa 8:22
136.4 C 222kPa 8:52
136.2 C 220kPa 9:22
136.4 C 218kPa 9:52
136.4 C 222kPa 10:23
136.3 C 222kPa 10:53
136.4 C 222kPa 11:22
VENTING 11:23
AIR DRYING 12:23
CYCLE COMPLETE 27:20

Digital Signature #
C02FDE173C2C2CD5
