
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002529
8:56 21/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:39
STERILIZING 3:45
136.4 C 224kPa 3:45
136.2 C 223kPa 4:15
136.2 C 223kPa 4:45
136.3 C 224kPa 5:15
136.4 C 224kPa 5:45
136.1 C 221kPa 6:15
136.2 C 222kPa 6:45
136.4 C 225kPa 7:15
136.4 C 225kPa 7:15
VENTING 7:15
AIR DRYING 8:08
CYCLE COMPLETE 21:15

Digital Signature #
CAA03CEC41F6839E
