
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001106
11:03 03/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:15
STERILIZING 3:20
136.0 C 221kPa 3:20
136.4 C 225kPa 3:50
136.4 C 222kPa 4:20
136.3 C 222kPa 4:50
136.4 C 223kPa 5:20
136.3 C 224kPa 5:50
136.4 C 225kPa 6:20
136.4 C 226kPa 6:50
136.4 C 226kPa 6:50
VENTING 6:50
AIR DRYING 7:44
CYCLE COMPLETE 16:07

Digital Signature #
CCEF44241DCB7F59
