
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000811
15:39 23/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 5:40
STERILIZING 7:23
136.4 C 223kPa 7:23
136.4 C 223kPa 7:53
136.4 C 222kPa 8:23
135.8 C 218kPa 8:53
136.3 C 222kPa 9:23
136.4 C 221kPa 9:53
136.3 C 223kPa 10:23
136.4 C 220kPa 10:53
136.4 C 220kPa 10:53
VENTING 10:53
AIR DRYING 11:56
CYCLE COMPLETE 12:18

Digital Signature #
D86BA0506938CC31
