
Midmark M11 - v1.0.0
Total Cycles: 3883

Name: RITTER 1
Sterilizer ID: V2178198
10/10/2023 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 193.1 F 4.4
4:00 213.5 F 9.7
6:00 224.1 F 6.5
8:00 235.4 F 11.0
10:00 234.5 F 8.4
12:00 240.0 F 10.3
14:00 247.2 F 13.9
16:00 259.9 F 20.8
18:00 271.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.2 F 28.8
2:00 271.9 F 28.9
3:00 272.1 F 28.6
4:00 272.1 F 29.0

Min 271.4 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 202.3 F 0.0
10:00 236.1 F 0.0
15:00 230.3 F 0.0
20:00 231.6 F 0.0
25:00 223.8 F 0.0
30:00 208.3 F 0.0
Drying Complete

FILLING: 2:36
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

10/10/2023 03:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____