
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002852
13:51 15/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:43
STERILIZING 3:39
136.4 C 223kPa 3:39
136.3 C 224kPa 4:09
136.4 C 224kPa 4:39
136.1 C 222kPa 5:09
136.1 C 219kPa 5:39
136.4 C 224kPa 6:09
136.4 C 224kPa 6:39
136.4 C 225kPa 7:09
136.4 C 225kPa 7:09
VENTING 7:09
AIR DRYING 8:02
CYCLE COMPLETE 8:03

Digital Signature #
006002E2E706C156
