
Midmark M11 - v1.0.5
Total Cycles: 2959

Name: RITTER2
Sterilizer ID: V2313228
10/02/2023 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.7 F 0.0
2:00 186.3 F 4.1
4:00 204.5 F 9.4
6:00 218.7 F 7.3
8:00 232.9 F 9.6
10:00 236.9 F 9.3
12:00 245.3 F 13.2
14:00 245.4 F 13.0
16:00 257.9 F 19.9
18:00 269.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.4 F 29.1
2:00 272.3 F 29.4
3:00 271.6 F 28.5
4:00 272.1 F 29.1

Min 271.4 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 228.1 F 0.0
10:00 217.7 F 0.0
15:00 208.4 F 0.0
20:00 210.4 F 0.0
25:00 182.8 F 0.0
30:00 177.6 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

10/02/2023 02:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____