
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004785
11:30 29/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:05
STERILIZING 4:08
136.4 C 221kPa 4:08
136.1 C 219kPa 4:38
136.1 C 223kPa 5:08
136.1 C 223kPa 5:38
136.4 C 224kPa 6:08
136.4 C 223kPa 6:38
136.0 C 220kPa 7:08
136.4 C 223kPa 7:38
136.4 C 223kPa 7:38
VENTING 7:39
AIR DRYING 8:30
CYCLE COMPLETE 12:28

Digital Signature #
3E79D2D5E1188A7A
