

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2994

Name: RITTER 1  
Sterilizer ID: V2178198  
03/13/2023 08:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.7 F 0.0  
2:00 143.0 F 0.0  
4:00 154.1 F 1.6  
6:00 187.3 F 4.5  
8:00 202.9 F 8.8  
10:00 219.5 F 13.5  
12:00 227.8 F 8.0  
14:00 239.5 F 12.9  
16:00 238.7 F 10.1  
18:00 236.6 F 8.4  
20:00 249.1 F 14.7  
22:00 260.9 F 21.3  
24:00 271.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.8  
2:00 271.5 F 28.3  
3:00 272.5 F 28.8  
4:00 271.5 F 28.2

Min 271.4 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 224.0 F 0.0  
10:00 235.3 F 0.0  
15:00 234.0 F 0.0  
20:00 218.9 F 0.0  
25:00 214.6 F 0.0  
30:00 211.8 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 24:35  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:02:35  
CYCLE COMPLETE

03/13/2023 10:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_