
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000481
14:31 05/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:30
PRESSURIZING 3:34
STERILIZING 4:53
136.0 C 229kPa 4:53
136.3 C 230kPa 5:23
136.3 C 231kPa 5:53
136.1 C 231kPa 6:23
136.2 C 230kPa 6:53
136.2 C 230kPa 7:23
136.4 C 233kPa 7:53
136.4 C 232kPa 8:23
136.4 C 232kPa 8:23
VENTING 8:24
AIR DRYING 9:18
CYCLE COMPLETE 11:31

Digital Signature #
E7AC22777BCDFA13
