

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4546

Name: RITTER 1  
Sterilizer ID: V2178198  
05/13/2024 09:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.2 F 0.0  
2:00 171.8 F 2.1  
4:00 169.8 F 5.3  
6:00 182.4 F 8.7  
8:00 209.5 F 12.6  
10:00 213.1 F 6.0  
12:00 222.7 F 9.9  
14:00 235.0 F 9.9  
16:00 235.8 F 10.1  
18:00 237.2 F 8.9  
20:00 245.7 F 13.4  
22:00 256.4 F 19.1  
24:00 266.3 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.4 F 29.4  
2:00 272.5 F 29.2  
3:00 271.5 F 28.7  
4:00 271.2 F 28.7

Min 271.1 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.0 F 0.0  
5:00 215.8 F 0.0  
10:00 210.4 F 0.0  
15:00 213.3 F 0.0  
20:00 209.9 F 0.0  
25:00 209.7 F 0.0  
30:00 208.6 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 25:51  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:04:42  
CYCLE COMPLETE

05/13/2024 11:02 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_