
Midmark M11 - v1.0.5
Total Cycles: 418

Name: RITTER2
Sterilizer ID: V2313228
10/05/2021 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.4 F 0.0
2:00 157.0 F 2.2
4:00 166.3 F 5.6
6:00 188.7 F 10.3
8:00 215.6 F 5.7
10:00 226.8 F 10.8
12:00 229.1 F 7.4
14:00 242.5 F 13.4
16:00 242.3 F 12.0
18:00 254.3 F 18.3
20:00 265.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 271.5 F 29.4
2:00 271.6 F 29.7
3:00 271.7 F 29.8
4:00 271.6 F 29.9

Min 271.3 F 29.1
Max 272.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 220.2 F 0.0
10:00 224.0 F 0.0
15:00 227.8 F 0.0
20:00 223.7 F 0.0
25:00 228.4 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:42
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

10/05/2021 12:26 PM
ACCEPTED: X
REJECTED:
Operator: U01