
Midmark M11 - v1.0.5
Total Cycles: 1749

Name: RITTER2
Sterilizer ID: V2313228
11/21/2022 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	127.9 F	0.0
2:00	127.4 F	0.0
4:00	139.8 F	0.1
6:00	182.4 F	0.6
8:00	203.9 F	3.1
10:00	224.4 F	7.6
12:00	236.0 F	12.5
14:00	235.8 F	9.8
16:00	238.8 F	9.1
18:00	241.5 F	10.9
20:00	248.2 F	14.5
22:00	260.7 F	21.5
24:00	272.0 F	29.0

STERILIZING

mm:ss	Degrees	PSI
0:00	271.3 F	28.7
1:00	272.6 F	29.4
2:00	271.8 F	28.7
3:00	271.9 F	29.3
4:00	272.4 F	29.3

Min 271.3 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

mm:ss	Degrees	PSI
0:00	207.6 F	0.0
5:00	216.4 F	0.0
10:00	228.2 F	0.0
15:00	230.7 F	0.0
20:00	227.7 F	0.0
25:00	227.9 F	0.0
30:00	226.4 F	0.0

Drying Complete

DRYING
Drying Start
FILLING: 0:56
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

11/21/2022 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____