

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1893

Name: RITTER2  
Sterilizer ID: V2313228  
01/04/2023 12:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
2:00 179.4 F 3.6  
4:00 196.9 F 8.5  
6:00 216.4 F 9.2  
8:00 229.0 F 9.5  
10:00 233.7 F 8.9  
12:00 242.7 F 12.3  
14:00 245.4 F 13.5  
16:00 259.2 F 20.8  
18:00 271.0 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.1 F 29.3  
2:00 271.8 F 29.5  
3:00 272.0 F 29.2  
4:00 271.6 F 29.3  
  
Min 271.3 F 28.8  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.4 F 0.0  
5:00 234.8 F 0.0  
10:00 231.2 F 0.0  
15:00 226.6 F 0.0  
20:00 235.9 F 0.0  
25:00 230.9 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 1:49  
HEATING: 18:42  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:57:37  
CYCLE COMPLETE

01/04/2023 01:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_