
Midmark M11 - v1.0.0
Total Cycles: 2078

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2022 05:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.0 F 0.0
2:00 205.0 F 3.7
4:00 218.9 F 8.7
6:00 234.4 F 9.4
8:00 237.4 F 9.9
10:00 235.6 F 7.8
12:00 244.8 F 13.3
14:00 244.2 F 12.0
16:00 257.1 F 18.9
18:00 268.5 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.2
2:00 272.7 F 28.7
3:00 271.6 F 28.2
4:00 271.9 F 28.9

Min 271.4 F 28.1
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 230.9 F 0.0
10:00 235.7 F 0.0
15:00 230.4 F 0.0
20:00 229.5 F 0.0
25:00 228.3 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

06/29/2022 06:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____