
Midmark M11 - v1.0.5
Total Cycles: 4314

Name: RITTER2
Sterilizer ID: V2313228
01/29/2025 05:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 166.4 F 3.1
4:00 180.5 F 7.5
6:00 202.1 F 12.2
8:00 225.8 F 6.4
10:00 234.1 F 11.3
12:00 233.1 F 8.5
14:00 241.6 F 11.3
16:00 245.1 F 13.4
18:00 257.1 F 19.8
20:00 267.9 F 26.7
22:00 271.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 273.2 F 30.0
2:00 271.4 F 29.9
3:00 272.6 F 29.9
4:00 271.3 F 29.8

Min 271.1 F 29.5
Max 273.3 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 235.0 F 0.0
10:00 231.7 F 0.0
15:00 190.1 F 0.0
20:00 177.7 F 0.0
25:00 178.7 F 0.0
30:00 175.3 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

01/29/2025 06:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____