
Midmark M11 - v1.0.0
Total Cycles: 1726

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2022 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.5 F 0.0
2:00 186.9 F 3.5
4:00 200.8 F 8.0
6:00 221.1 F 9.2
8:00 230.5 F 9.0
10:00 239.0 F 9.9
12:00 242.6 F 12.3
14:00 241.6 F 11.1
16:00 255.2 F 18.1
18:00 267.1 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.0 F 28.5
2:00 272.5 F 29.0
3:00 271.5 F 28.6
4:00 272.2 F 28.7

Min 271.4 F 28.3
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 233.5 F 0.0
10:00 231.7 F 0.0
15:00 231.0 F 0.0
20:00 220.6 F 0.0
25:00 215.9 F 0.0
30:00 215.3 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:26
CYCLE COMPLETE

03/16/2022 04:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____