

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4412

Name: RITTER2  
Sterilizer ID: V2313228  
03/07/2025 02:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.6 F 0.0  
2:00 110.4 F 0.8  
4:00 115.0 F 3.0  
6:00 148.1 F 6.3  
8:00 169.2 F 10.1  
10:00 195.6 F 6.5  
12:00 224.3 F 6.5  
14:00 233.7 F 10.6  
16:00 232.3 F 8.0  
18:00 241.5 F 12.3  
20:00 239.5 F 10.7  
22:00 251.5 F 16.8  
24:00 262.2 F 23.0  
26:00 272.0 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.8  
1:00 271.8 F 29.8  
2:00 272.8 F 30.1  
3:00 272.1 F 30.4  
4:00 273.8 F 31.0

Min 270.8 F 29.7  
Max 274.3 F 31.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
5:00 236.6 F 0.0  
10:00 230.6 F 0.0  
15:00 223.6 F 0.0  
20:00 229.0 F 0.0  
25:00 224.6 F 0.0  
30:00 228.1 F 0.0

Drying Complete

FILLING: 0:57  
HEATING: 27:16  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:06:06  
CYCLE COMPLETE

03/07/2025 03:26 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_