
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001126
13:09 20/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 4:26
STERILIZING 5:54
136.4 C 225kPa 5:54
136.4 C 225kPa 6:24
136.4 C 225kPa 6:54
136.4 C 225kPa 7:24
136.1 C 226kPa 7:54
136.3 C 225kPa 8:24
136.4 C 226kPa 8:54
136.4 C 226kPa 9:24
136.4 C 226kPa 9:24
VENTING 9:24
AIR DRYING 10:22
CYCLE COMPLETE 25:22

Digital Signature #
F7D61AE243BD8521
