
Midmark M11 - v1.0.5
Total Cycles: 2153

Name: RITTER2
Sterilizer ID: V2313228
03/16/2023 07:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.9 F 0.0
2:00 79.1 F 0.0
4:00 89.0 F 0.0
6:00 118.9 F 1.3
8:00 148.2 F 4.1
10:00 173.0 F 7.7
12:00 195.7 F 11.6
14:00 218.5 F 4.7
16:00 218.8 F 8.6
18:00 230.6 F 12.7
20:00 233.8 F 8.8
22:00 240.8 F 11.9
24:00 242.7 F 12.1
26:00 254.0 F 17.9
28:00 264.4 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 271.9 F 29.0
2:00 272.0 F 29.4
3:00 271.5 F 29.5
4:00 271.2 F 28.9

Min 271.0 F 28.8
Max 272.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 232.5 F 0.0
10:00 188.5 F 0.0
15:00 179.5 F 0.0
20:00 176.7 F 0.0
25:00 176.1 F 0.0
30:00 175.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:60
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:07:58
CYCLE COMPLETE

03/16/2023 08:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____