
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001453
9:07 14/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:35
STERILIZING 3:39
136.0 C 220kPa 3:39
136.4 C 224kPa 4:09
136.4 C 224kPa 4:39
136.4 C 224kPa 5:09
136.4 C 223kPa 5:39
136.2 C 224kPa 6:09
136.2 C 221kPa 6:39
136.3 C 223kPa 7:09
136.3 C 223kPa 7:09
VENTING 7:09
AIR DRYING 8:01
CYCLE COMPLETE 8:08

Digital Signature #
D978A311F81CC13D
