
Midmark M11 - v1.0.5
Total Cycles: 1365

Name: RITTER2
Sterilizer ID: V2313228
08/03/2022 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 186.8 F 0.4
4:00 198.2 F 4.7
6:00 214.8 F 9.6
8:00 225.6 F 8.7
10:00 235.3 F 11.4
12:00 236.2 F 9.5
14:00 237.6 F 9.4
16:00 249.5 F 15.3
18:00 261.5 F 22.0
20:00 271.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.3 F 29.3
2:00 271.5 F 29.0
3:00 272.3 F 29.2
4:00 271.4 F 28.8

Min 271.3 F 28.6
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 235.2 F 0.0
10:00 223.9 F 0.0
15:00 232.7 F 0.0
20:00 228.3 F 0.0
25:00 230.7 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

08/03/2022 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____