
Midmark M11 - v1.0.5
Total Cycles: 3638

Name: RITTER2
Sterilizer ID: V2313228
05/23/2024 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.3 F 0.0
2:00 123.9 F 1.1
4:00 136.4 F 3.3
6:00 164.9 F 6.7
8:00 187.8 F 10.6
10:00 205.0 F 4.9
12:00 217.1 F 6.4
14:00 228.9 F 10.5
16:00 232.8 F 8.4
18:00 241.7 F 12.5
20:00 240.9 F 11.2
22:00 252.7 F 17.2
24:00 263.2 F 23.5
26:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.8 F 29.7
2:00 272.0 F 29.0
3:00 272.1 F 29.6
4:00 272.7 F 29.6

Min 271.6 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.7 F 0.0
5:00 207.2 F 0.0
10:00 194.2 F 0.0
15:00 186.8 F 0.0
20:00 180.7 F 0.0
25:00 176.7 F 0.0
30:00 176.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:21
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:04:48
CYCLE COMPLETE

05/23/2024 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____