
Midmark M11 - v1.0.5
Total Cycles: 490

Name: RITTER2
Sterilizer ID: V2313228
10/27/2021 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 185.1 F 3.2
4:00 197.4 F 8.2
6:00 216.3 F 13.4
8:00 228.3 F 8.8
10:00 237.0 F 10.7
12:00 242.8 F 12.7
14:00 243.1 F 12.2
16:00 255.6 F 18.8
18:00 266.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.6
2:00 271.6 F 28.9
3:00 272.7 F 29.8
4:00 272.3 F 29.2

Min 271.5 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 225.6 F 0.0
10:00 213.4 F 0.0
15:00 196.4 F 0.0
20:00 195.8 F 0.0
25:00 203.6 F 0.0
30:00 205.0 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

10/27/2021 11:46 AM
ACCEPTED: X
REJECTED:
Operator: U01