
Midmark M11 - v1.0.5
Total Cycles: 2150

Name: RITTER2
Sterilizer ID: V2313228
03/15/2023 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 187.7 F 3.6
4:00 205.3 F 9.2
6:00 218.7 F 7.5
8:00 232.2 F 11.0
10:00 234.9 F 9.0
12:00 237.9 F 9.6
14:00 249.6 F 15.5
16:00 262.3 F 22.8
18:00 272.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.8 F 29.6
2:00 271.5 F 28.8
3:00 272.5 F 29.8
4:00 272.2 F 29.3

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 214.4 F 0.0
10:00 227.8 F 0.0
15:00 223.4 F 0.0
20:00 223.9 F 0.0
25:00 229.3 F 0.0
30:00 233.7 F 0.0

Drying Complete

FILLING: 2:01
HEATING: 18:13
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:28
CYCLE COMPLETE

03/15/2023 11:02 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____