
Midmark M11 - v1.0.5
Total Cycles: 4283

Name: RITTER2
Sterilizer ID: V2313228
01/22/2025 07:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.7 F 0.0
2:00 152.3 F 2.5
4:00 168.0 F 6.2
6:00 191.6 F 10.8
8:00 220.5 F 4.7
10:00 231.5 F 9.5
12:00 235.9 F 10.0
14:00 241.5 F 11.8
16:00 239.5 F 10.4
18:00 252.9 F 17.0
20:00 264.4 F 23.9
22:00 270.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 271.7 F 29.4
2:00 271.6 F 29.5
3:00 271.6 F 29.4
4:00 272.0 F 29.3

Min 271.6 F 29.3
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 231.8 F 0.0
10:00 185.9 F 0.0
15:00 180.5 F 0.0
20:00 178.2 F 0.0
25:00 177.1 F 0.0
30:00 173.8 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:41
CYCLE COMPLETE

01/22/2025 08:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____