

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002627  
7:27 28/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:57  
PRESSURIZING 3:00  
STERILIZING 4:11  
136.4 C 224kPa 4:11  
136.4 C 225kPa 4:41  
136.3 C 223kPa 5:11  
136.4 C 224kPa 5:41  
136.1 C 222kPa 6:11  
136.1 C 221kPa 6:41  
136.1 C 221kPa 7:11  
136.4 C 222kPa 7:41  
136.4 C 222kPa 7:41  
VENTING 7:41  
AIR DRYING 8:32  
CYCLE COMPLETE 12:59

-----  
Digital Signature #  
42C656BA72C1D337  
-----