
Midmark M11 - v1.0.0
Total Cycles: 5097

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2024 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.5 F 0.0
2:00 161.9 F 0.2
4:00 177.1 F 0.6
6:00 199.2 F 1.0
8:00 214.0 F 4.5
10:00 224.5 F 8.3
12:00 235.5 F 12.5
14:00 234.2 F 8.8
16:00 243.5 F 11.6
18:00 244.6 F 12.5
20:00 243.1 F 11.6
22:00 255.4 F 17.9
24:00 266.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.3 F 28.8
2:00 272.4 F 28.4
3:00 271.6 F 28.0
4:00 271.6 F 28.4

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 205.6 F 0.0
10:00 189.7 F 0.0
15:00 190.2 F 0.0
20:00 204.7 F 0.0
25:00 214.0 F 0.0
30:00 215.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:49
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:04:37
CYCLE COMPLETE

11/15/2024 12:24 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____