
Midmark M11 - v1.0.5
Total Cycles: 3774

Name: RITTER2
Sterilizer ID: V2313228
07/12/2024 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 71.0 F 0.0
4:00 78.0 F 0.0
6:00 95.9 F 0.1
8:00 140.5 F 0.4
10:00 173.5 F 3.1
12:00 196.1 F 6.4
14:00 200.7 F 9.8
16:00 213.1 F 13.1
18:00 220.9 F 6.4
20:00 229.7 F 10.4
22:00 233.6 F 9.1
24:00 241.0 F 11.9
26:00 238.5 F 10.1
28:00 250.5 F 15.9
30:00 260.7 F 21.8
32:00 270.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.4 F 29.6
2:00 272.1 F 28.9
3:00 272.2 F 29.7
4:00 272.2 F 29.1

Min 271.4 F 28.8
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 201.7 F 0.0
10:00 183.1 F 0.0
15:00 178.5 F 0.0
20:00 177.8 F 0.0
25:00 176.9 F 0.0
30:00 178.5 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 32:45
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:11:09
CYCLE COMPLETE

07/12/2024 09:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____