
Midmark M11 - v1.0.5
Total Cycles: 948

Name: RITTER2
Sterilizer ID: V2313228
04/04/2022 07:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 89.1 F 0.0
4:00 105.8 F 0.0
6:00 98.8 F 0.0
8:00 154.2 F 0.2
10:00 197.2 F 0.3
12:00 207.5 F 1.1
14:00 219.8 F 4.6
16:00 228.9 F 7.9
18:00 238.0 F 11.7
20:00 234.2 F 8.4
22:00 244.5 F 12.2
24:00 244.4 F 12.7
26:00 243.5 F 12.1
28:00 256.2 F 18.8
30:00 267.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.8 F 29.2
2:00 271.7 F 28.5
3:00 272.8 F 29.3
4:00 271.9 F 28.5

Min 271.6 F 28.4
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 229.2 F 0.0
10:00 180.7 F 0.0
15:00 171.5 F 0.0
20:00 159.1 F 0.0
25:00 158.2 F 0.0
30:00 151.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:26
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:20
CYCLE COMPLETE

04/04/2022 08:51 AM
ACCEPTED: X
REJECTED:
Operator: U02