
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008486
10:07 28/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:31
STERILIZING 3:39
136.4 C 223kPa 3:39
135.9 C 221kPa 4:09
136.1 C 220kPa 4:39
136.0 C 222kPa 5:09
136.2 C 218kPa 5:39
136.4 C 221kPa 6:09
136.4 C 223kPa 6:39
136.4 C 223kPa 7:09
136.4 C 223kPa 7:09
VENTING 7:09
AIR DRYING 7:58
CYCLE COMPLETE 37:58

Digital Signature #
5B570666E2F015D6
