
Midmark M11 - v1.0.0
Total Cycles: 2880

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2023 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.2 F 0.0
2:00 196.2 F 4.0
4:00 199.6 F 8.7
6:00 221.1 F 8.8
8:00 228.7 F 9.2
10:00 239.0 F 9.3
12:00 241.4 F 12.3
14:00 240.8 F 10.8
16:00 254.3 F 17.7
18:00 266.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.1 F 28.8
2:00 272.2 F 29.1
3:00 271.7 F 28.5
4:00 272.4 F 29.2

Min 271.5 F 28.4
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 234.5 F 0.0
10:00 227.8 F 0.0
15:00 231.3 F 0.0
20:00 227.5 F 0.0
25:00 227.8 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

02/09/2023 11:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____